


Certificate of Analysis Micropipettes - Biopsy

Batchnumber	20251104
Production date	04 November 2025
Expiry date	31 October 2030

Article No.

BP-25-35°-bl
BP-30-30°-be
BP-35-30°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	86	≥ 80 %

Release date	15 January 2026
Released by	Nicole Geiger
Signature	

REV 04_00 / 2023-09-15