

Certificate of Analysis Micropipettes - Holding

Batchnumber	20251104
Production date	04 November 2025
Expiry date	31 October 2030

Article No.

HP-80-30°
HP-80-35°

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	86	≥ 80 %

Release date	15 January 2026
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 04_00 / 2023-09-15