

**Certificate of Analysis
Micropipettes - Biopsy**

Batchnumber 20251223
 Production date 23 December 2025
 Expiry date 31 October 2030

Article No.
 BP-35-35°-bl

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 25 February 2026
 Released by Dr. René Richarz
 Signature



REV 04_00 / 2023-09-15