

**Certificate of Analysis  
Micropipettes - Holding**

Batchnumber 20251223  
 Production date 23 December 2025  
 Expiry date 31 October 2030

**Article No.**

HP-100-30°  
 HP-110-30°

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 25 February 2026  
 Released by Dr. René Richarz  
 Signature



REV 04\_00 / 2023-09-15