

Certificate of Analysis Micropipettes - Holding

Batchnumber 20260218

Production date 18 February 2026

Expiry date 28 February 2031

Article No.

HP-100-35°

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 23 March 2026

Released by Dr. René Richarz

Signature



REV 04_00 / 2023-09-15