


Certificate of Analysis Micropipettes - ICSI

Batchnumber	20260409	
Production date	09 April 2026	
Expiry date	25 March 2031	

Article No.

IC-5-30°-s
IC-5-35°-s
IC-SL-4.6-30°-s

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	01 June 2026	
Released by	Nicole Geiger	
Signature		

REV 05_00 / 2023-09-15