

Certificate of Analysis Micropipettes - ICSI

Batchnumber		20260713
Production date		13 July 2026
Expiry date		31 July 2031

Article No.

IC-5-35°-s

Parameter	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date		09 September 2026
--------------	--	-------------------

Released by		Nicole Geiger
-------------	--	---------------

Signature

pp. P. Geiger

REV 05_00 / 2023-09-15