

**Certificate of Analysis  
GM501 Collagenase**

Batchnumber	COLL-2102
Production date	22 January 2021
Expiry date	31 October 2021
Total volume (mL)	840

	<b>Results</b>	<b>Specifications</b>
pH	7,42	7,20-7,70
Osmolality	279	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	99	≥ 80 %

Release date	08 February 2021
Released by	Vanessa Bierlich
Signature	

P.P. 

REV 09\_00 /2021-02-08