



Kitazato Corporation

<http://www.kitazato.co.jp/>

PRODUCT: Hypure® Oil Light

LOT#: E0301

REF: 96004
Code: MOLHD-500

EXPIRY DATE : 2025/09/30
Year/Month/Day


APPROVED BY : Kazumi Shibata
Quality Control

DATE : 2024. 4. 16
Year/Month/Day

CERTIFICATE OF ANALYSIS

The product has been manufactured by Kitazato Corporation in conformity with requirements of ISO13485 Quality management system. The quality of products was tested and evaluated by quality control procedures

ASSAY	SPECIFICATION	RESULT
Color:	Clear, Particle Free	Clear Particle Free
Endotoxin:	< 0,1 EU/ml	<0.1
Sterility:	No microbial Growth	No growth
MEA 1:		
- Control:	≥ 80%	100
- This Lot MEA 1:	≥ 80%	95.24
- This Lot MEA 2:	≥ 80%	100
POV:	<0,1 mEq/Kg	<0.06

¹ One cell mouse embryo testing is performed using fresh mouse embryos. Test results indicate the percentage of mouse embryos developing to fully expanded blastocysts after 120 hours in culture medium.