

Certificate of Analysis Oocyte Aspiration Needle

Product code (REF)	GY-711924LF
Batchnumber	712303/36/03
Production Date	6. September 2023
Expiry date	31. August 2026

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	24 October 2023
Released by	Nicole Geiger
Signature	

pp. N. Geiger