

**Certificate of Analysis
Oocyte Aspiration Needle**

Product code (REF)	GY-711724LF
Batchnumber	712403/27/01
Production Date	1. Juli 2024
Expiry date	30. Juni 2027

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 30 August 2024

Released by Nicole Geiger

Signature

