

**Certificate of Analysis  
Oocyte Aspiration Needle**

Product code (REF)	GY-711724LF
Batchnumber	712403/38/01
Production Date	16. September 2024
Expiry date	31. August 2027

	<b>Result</b>	<b>Specification</b>
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 30 October 2024

Released by Joana Sturm

Signature

