


**Certificate of Analysis
Oocyte Aspiration Needle**

Product code (REF) GY-711724LF
 Batchnumber 712404/43/01
 Expiry date 30. September 2027

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 02 January 2025
 Released by Nicole Geiger
 Signature



REV 03_00 / 2023-09-15