

**Certificate of Analysis
Oocyte Aspiration Needle**

Batchnumber	712502/26/01
Production date	23 June 2025
Expiry date	31 May 2028

Article No.

GY-711924LF

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	05 August 2025
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 04_00 / 2025-05-21