

**Certificate of Analysis
Oocyte Aspiration Needle**

Batchnumber	712503/31/01
Production date	01 August 2025
Expiry date	31 July 2028

Article No.

GY-711724LF

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	09 September 2025
Released by	Nicole Geiger
Signature	

pp. P. Geiger

REV 04_00 / 2025-05-21