

Certificate of Analysis Oocyte Aspiration Needle

Batchnumber	712504/47/01
Production date	17 November 2025
Expiry date	31 October 2028

Article No.

GY-711724LF

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date	04 February 2026
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 04_00 / 2025-05-21