

Certificate of Analysis Oocyte Aspiration Needle

Batchnumber	712504/49/01
Production date	01 December 2025
Expiry date	30 November 2028

Article No.

GY-711724LF

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	21 January 2026
Released by	René Richarz
Signature	



REV 04_00 / 2025-05-21