

**Certificate of Analysis  
Oocyte Aspiration Needle**

Batchnumber 712601/03/01  
 Production date 12 January 2026  
 Expiry date 31 December 2028

**Article No.**

GY-711724LF-A

<b>Parameter</b>	<b>Result</b>	<b>Specification</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 25 February 2026  
 Released by Dr. René Richarz  
 Signature



REV 04\_00 / 2025-05-21