

**Certificate of Analysis
Oocyte Aspiration Needle**

Product code (REF)	GY-721724LF
Batchnumber	722404/44/01
Production Date	28 October 2024
Expiry date	30 September 2027

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 16 December 2024

Released by Nicole Geiger

Signature

