

**Certificate of Analysis  
Oocyte Aspiration Needle**

|                    |                      |
|--------------------|----------------------|
| Product code (REF) | <b>GY-93xxxxxx</b>   |
| Batchnumber        | <b>93220201</b>      |
| Production Date    | <b>12 April 2022</b> |
| Expiry date        | <b>30 April 2025</b> |

|   | <b>Result</b> | <b>Specification</b> |
|---|---------------|----------------------|
| <b>Mouse Embryo Assay</b><br>(1-cell, blastocysts after 96 hours) | 90,48         | ≥ 80 %               |

|              |                     |
|--------------|---------------------|
| Release date | 23 May 2022         |
| Released by  | Dr. Julia Heinzmann |
| Signature    |                     |

*p.p. Julia Heinzmann*