

## Certificate of Analysis Oocyte Aspiration Needle

Product code (REF)	<b>GY-931667LF-A</b>
Batchnumber	<b>93220301</b>
Production Date	<b>21 July 2022</b>
Expiry date	<b>31 July 2025</b>

	<b>Result</b>	<b>Specification</b>
<b>Mouse Embryo Assay</b> (1-cell, blastocysts after 96 hours)	95,24	≥ 80 %

Release date	02 September 2022
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*