

**Certificate of Analysis
Oocyte Aspiration Needle**

Product code (REF)	GY-93220401
Batchnumber	93220401
Production Date	12 October 2022
Expiry date	31 October 2025

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90,48	≥ 80 %

Release date 19 January 2023

Released by Nicole Geiger

Signature

