

**Certificate of Analysis
Oocyte Aspiration Needle**

Product code (REF)	GY-931627LF-A
Batchnumber	932403/34/01
Production Date	20. August 2024
Expiry date	31. Juli 2027

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 06 November 2024

Released by Nicole Geiger

Signature

