

**Certificate of Analysis
Oocyte Aspiration Needle**

Product code (REF)	GY-931627LF-A
Batchnumber	932403/38/01
Production Date	16. September 2024
Expiry date	30. August 2027

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	86	≥ 80 %

Release date 27 November 2024

Released by Joana Sturm

Signature

