

Certificate of Analysis
Oocyte Aspiration Needle

Batchnumber	932502/15/01
Production date	10 April 2025
Expiry date	31 March 2028

Article No.

GY-931768LF

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	10 July 2025
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 04_00 / 2025-05-21