

CERTIFICATE OF QUALITY

| | |
|--------------------------------|--|
| ITEM name and code | #3000 Strip-ID 200/10 - Denuding Pipette |
| LOT NUMBER | 20240305 |
| EXPIRY DATE | 5 March 2028 |
| STERILIZATION DATE | 8 March 2024 |
| STERILIZATION TECHNIQUE | Gamma Irradiation |
| STERILIZATION DOSE | 22,5 - 45 kGy |

QUALITY CONTROL TESTING PROCEDURES

| TEST | SPECIFICATION | CONCLUSION |
|---|----------------|------------|
| Endotoxin Assay (LAL) | <1.0 EU/device | Pass * |
| TEST | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA) | > 80% | Pass * |
| <i>expressed as day 5 expanded blastocyst stage</i> | | |

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Merve Toprak
Gynetics Medical Products N.V.

Signature date:
4 April 2024