

CERTIFICATE OF QUALITY

ITEM name and code	#3000 Strip-ID 135/10 - Denuding Pipette
LOT NUMBER	20250105
EXPIRY DATE	5 January 2029
STERILIZATION DATE	10 January 2025
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGy

QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA)	> 80%	Pass *
<i>expressed as day 5 expanded blastocyst stage</i>		

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Eef Wouters
Gynetics Medical Products N.V.

Signature date:
4 February 2025