

CERTIFICATE OF ANALYSIS

Product	20 µm chamber depth - 2 analysis chambers	
Slide reference	SC 20-01-02-B-CE	
Batch Machine	082551	L
Production date range (DD-MM-YYYY)	21-02-2025 - 21-02-2025	
Best before (Month-YY)	February-26	

Analysis	Criteria	Result
Dimensions and microscope inspection	20 inspection points	PASSED
Semen Toxicity Test (internal protocol): <i>Hamilton Thorne IVOS II Digital</i> Type of semen: <i>Porcine</i>	Time degradation 4 minutes <8%	PASSED
Chamber height: <i>Low Coherence Interferometry</i>	Median <10% from indicated mean. Variation <2%	PASSED
Average chamber height (µm)		20,15

**WE CERTIFY THAT THE ABOVE MENTIONED PRODUCTS
ARE IN COMPLIANCE WITH OUR SPECIFICATIONS**

Date	February 21, 2025
Name	Anne Linda van Kappel
Qualification	General Manager
Signature	