

## Mouse Embryo Assay Certificate of Analysis

Manufacturer		Microtech
Product		Pipette
Batchnumber		2504021
Production date		02 April 2025
Expiry date		30 April 2029
Mouse Embryo Test	<b><u>Result</u></b>	<b><u>Specifications</u></b>
	100	≥ 80 %

### **Control assay method and results:**

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium:

20/21 (95 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

*For a valid assay, Embryotech<sup>TM</sup> requires at least 80 % of one cell stage control embryos to develop to expanded blastocyst within 96-hours.*

### **Test assay method and results:**

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium that was extracted from the test article:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

Release date	21 May 2025
Released by	Nicole Geiger
Signature	

pp. N. Geiger

Rev04\_00 / 2023-09-15