

Mouse Embryo Assay Certificate of Analysis

Manufacturer	Microtech	
Product	Micropipette	
Batchnumber	2512031	
Expiry date	31 December 2029	
Mouse Embryo Test	<u>Result</u>	<u>Specifications</u>
	95	≥ 80 %

Control assay method and results:

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

For a valid assay, Embryotech™ requires at least 80 % of one cell stage control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results:

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium that was extracted from the test article:

21/21 (100 %)	1-cell to 2-cell within 24 hr
20/21 (95 %)	1-cell to expanded blastocyst within 96 hr

Release date	21 January 2026
Released by	René Richarz
Signature	