

Mouse Embryo Assay Certificate of Analysis

Manufacturer		Microtech
Product		Micropipettes
Batchnumber		2601211
Expiry date		31 January 2030
Mouse Embryo Test	<u>Result</u>	<u>Specifications</u>
	100	≥ 80 %

Control assay method and results:

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

For a valid assay, EmbryotechTM requires at least 80 % of one cell stage control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results:

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium that was extracted from the test article:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

Release date	25 February 2026
Released by	Dr. René Richarz
Signature	



Rev04_00 / 2023-09-15