

**Mouse Embryo Assay  
Certificate of Analysis**

Manufacturer		Microtech
Product		Micropipette
Batchnumber		2604011
Expiry date		30 April 2030
Mouse Embryo Test	<b><u>Result</u></b>	<b><u>Specifications</u></b>
	100	≥ 80 %

**Control assay method and results:**

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

*For a valid assay, Embryotech™ requires at least 80 % of one cell stage control embryos to develop to expanded blastocyst within 96-hours.*

**Test assay method and results:**

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium that was extracted from the test article:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

Release date	19 May 2026
Released by	Dr. René Richarz
Signature	



Rev04\_00 / 2023-09-15