Mouse Embryo Assay
Certificate of Analysis

Manufacturer		Microtech
Product		Micropipettes
Batchnumber		2607291
Expiry date		31 July 2030
Mouse Embryo Test	***Result***	***Specifications***
	100	**≥ 80 %**

Control assay method and results:

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

For a valid assay, Embryotech™ requires at least 80 % of one cell stage control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results:

21 one cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium that was extracted from the test article:

21/21 (100 %)	1-cell to 2-cell within 24 hr
21/21 (100 %)	1-cell to expanded blastocyst within 96 hr

Release date	29 October 2026
Released by	Dr. René Richarz
Signature	

Gynemed GmbH Co. KG, Wagrienring 24b, 23730 Sierksdorf, Germany
phone: +49-4561-5132-500, fax: +49-4561-5132-519, e-mail: info@gynemed.de