

CERTIFICATE OF ANALYSIS

Product	20 µm chamber depth - 4 analysis chambers	
Slide reference	SC 20-01-04-B-CE	
Batch Machine	452452	L
Production date range (DD-MM-YYYY)	05-11-2024 - 05-11-2024	
Best before (Month-YY)	November-25	

Analysis	Criteria	Result
Dimensions and microscope inspection	20 inspection points	PASSED
Semen Toxicity Test (internal protocol): <i>Hamilton Thorne IVOS II Digital</i>	Time degradation 4 minutes <8%	PASSED
Type of semen: <i>Porcine</i>		
Chamber height: <i>Low Coherence Interferometry</i>	Median <10% from indicated mean. Variation <2%	PASSED
Average chamber height (µm)		20,24

**WE CERTIFY THAT THE ABOVE MENTIONED PRODUCTS
ARE IN COMPLIANCE WITH OUR SPECIFICATIONS**

Date	November 5, 2024
Name	Anne Linda van Kappel
Qualification	General Manager
Signature	