



We care about your needs

Halotech DNA SL.
C/ Faraday 7 , 28049 Madrid
Spain
Tel: +34 91 279 6950
info@halotech.es
www.halotechdna.com

CERTIFICATE OF ANALYSIS v1.0

Product name: Ovoselect®
Product number: HT-OVS16
Lot Number: OVS16R 2402

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Lysis Solution, pH 7,5±1, Colorless, Transparent, 10ml.

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Stabilization Solution, pH 7,0±0,5, Odorless, Transparent, 10 ml.

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Supercoated slides, Translucent appearance, Homogeneity, surface completely covered, Surface completely covered.

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Eppendor tube with Agarose, Color, State, Aliquoted volume, Colorless, Solid, 0,5 ml.

Results with biological sample, comparison two tests: Accepted/Rejected (AC/RE)

Table with 2 columns: Expiration information, Production Date, Expiration Date, Signature Date. Values: 27/11/2024, 27/05/2025, 28/11/2024.

Handwritten signature of Jaime Gosálvez
Jaime Gosálvez
Technical Director

NOTICE: HalotechDNA products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from HalotechDNA.